




EIE INSTRUMENTS PVT. LTD.
(An ISO 9001:2015 Certified Company)

ACRYLIC - SLUMP TEST APPARATUS - MINI



ACRYLIC - SLUMP TEST APPARATUS - MINI

- **Made of Acrylic**

- The mini-slump test was used to find the spread of grout mixtures.
- Dimensions of mini-slump apparatus 57 mm high with at the top diameters of 19 mm and 38 mm at the bottom.
- It is placed on a flat acrylic or glass plate. (at extra cost)
- CBG mix is poured into a cone until it's full.
- Then the mixture is allowed to spread by removing the mini slump cone in vertical direction
- The spread diameters are measured and the average spread diameter is calculated.

Corporate Office:

A-1301, BVR Ek, Opp. Hotel Inder Residency,
Near Gujarat College, Ellisbridge,
Ahmedabad-380 006.
Phone : (079) 6621 1234
Website : www.eieinstruments.com

Calibration Laboratory:

Survey No. 1098, Nr. Mahakali Temple,
Opp. Primary School, Village Kubadthal,
Tal-Daskroi, Ahmedabad-382 433.
Phone: (079)-3520 8300
E-mail : info@eieinstruments.com

GST TIN No. : 24AABCE4018L1ZU
I.T. PAN No. : AABCE 4018L
IEC No. : 0813026598 Dt. 04/03/2014
CIN No. : U29199GJ2004PTC045078
TAN No. : AHME00572F