



EIE INSTRUMENTS PVT. LTD.
(An ISO 9001:2015 Certified Company)

IS 2720, 9669 - ELECTRIC OPERATED LABORATORY CBR APPARATUS - WITH 25 KN PROVING RING



Complies with following testing standards

IS: 2720 (Part XVI), IS: 9669

Applications

EIE Motorized California Bearing Ratio Test Apparatus determines the load carrying capacity in the laboratory when in place density and water contents are such that the degree of saturation is 80 % or greater, when material is cohesionless and coarse grained.

Supplied complete with following accessories: 1 No each:

- **Load frame, motorized 5000 Kg (50 kn) capacity, with single rate of travel 1.25mm/min**
- C.B.R. Mould, 150 mm diameter x 175 mm high with extension collar & perforated base plate. Made of Mild steel.
- Penetration Piston, face diameter 50 mm with adjustable bracket.
- Perforated plate, 148 mm diameter with adjustable stem and lock nut-gunmetal
- Displacer block, 148 mm diameter x 47.7 mm high with detachable handle.
- Annular metal weight, 2.5 kg, 147 mm diameter with 53 mm diameter central hole.
- Slotted metal weight, 2.5 kg, 147 mm diameter with 53 mm diameter slot.
- Metal tripod stand for dial gauges.
- Rammer 2.6 kg weight with a drop of 310 mm
- Rammer 4.89 kg with weight with a drop of 450 mm
- **High sensitivity proving ring, 25 KN capacity fitted with 0.002 mm dial gauge in a nice teakwood case.**
- Dial gauge 0.01mm x 25mm / 30mm travel.

Click here for working video : <https://www.youtube.com/watch?v=hCWSTXztzSw>

Corporate Office:

A-1301, BVR Ek, Opp. Hotel Inder Residency,
Near Gujarat College, Ellisbridge,
Ahmedabad-380 006.
Phone : (079) 6621 1234
Website : www.eieinstruments.com

Calibration Laboratory:

Survey No. 1098, Nr. Mahakali Temple,
Opp. Primary School, Village Kubadthal,
Tal-Daskroi, Ahmedabad-382 433.
Phone: (079)-3520 8300
E-mail : info@eieinstruments.com

GST TIN No. : 24AABCE4018L1ZU
I.T. PAN No. : AABCE 4018L
IEC No. : 0813026598 Dt. 04/03/2014
CIN No. : U29199GJ2004PTC045078
TAN No. : AHME00572F