



EIE INSTRUMENTS PVT. LTD.
(An ISO 9001:2015 Certified Company)

MORTAR MIXER WITH DIGITAL RPM INDICATOR - 4.75 LTR



Compliance with following International Standards:

EN 196-1 & EN 196-3, IS : 4031, IS 10890, IS : 1727, ASTM C451

Construction details:

1. Compact and tactful design
 2. Aesthetic Appearance
 3. Exterior body powder coated in attractive shades
 4. Paddle and bowl made of Stainless Steel Material
 5. An epicyclic type stainless steel paddle imparts both planetary and revolving motion.
 6. Stainless steel bowl is fitted with handle - **Nominal capacity 4.75 Ltrs**
 7. Stainless Steel Handle is fitted with lid made of non-absorbing material and is attacked by cement, masonry cement, cement pozzolana mixture or lime pozzolana mixture.
-
1. Mixer operates on two different speeds. Mixer blade has low speed of $(140 \pm 5 \text{ rpm})$ and medium speed of $(285 \pm 10 \text{ rpm})$.
 2. It also has a planetary movement of $(62 \pm 5 \text{ rpm})$ in low range and planetary movement of $(125 \pm 10 \text{ rpm})$ in medium range.
 3. The mixing paddle has a planetary motion and is driven by a motor
 4. Suitable for operation on 230 Volts, 50 Hz, single phase, AC Supply
 5. No need to change gear
 6. Speed will be changed by pressing simple button
 7. Machine will display the running RPM
 8. Simple to operate

Corporate Office:

A-1301, BVR Ek, Opp. Hotel Inder Residency,
Near Gujarat College, Ellisbridge,
Ahmedabad-380 006.
Phone : (079) 6621 1234
Website : www.eieinstruments.com

Calibration Laboratory:

Survey No. 1098, Nr. Mahakali Temple,
Opp. Primary School, Village Kubadthal,
Tal-Daskroi, Ahmedabad-382 433.
Phone: (079)-3520 8300
E-mail : info@eieinstruments.com

GST TIN No. : 24AABCE4018L1ZU
I.T. PAN No. : AABCE 4018L
IEC No. : 0813026598 Dt. 04/03/2014
CIN No. : U29199GJ2004PTC045078
TAN No. : AHME00572F