



1. Available Variants

1. Rapid Moisture Meter – 0–25%
2. Rapid Moisture Meter – 0–50%

2. Variant Details

1. Rapid Moisture Meter – 0–25%

Compliance Standards

IS:2720 (PART II) / IS 12175, ASTM D4944, AASHTO T217, BS 812-109

Details

The unit consists of a pressure vessel with clamp for sealing cap, rubber sealing gasket, pressure gauge calibrated in percentage moisture content 0-25% X 1% on the wet weight basis, a counter poised balance for weighing sample, a scoop for measuring carbide reagent, a bottle of reagent, one cleaning brush and a set of 4 steel balls (3 OF 12.5 MM & 1 OF 25 MM DIA) for thorough mixing. Complete in highly polished wooden carrying case with handle.

Specifications

Specification	Details
Moisture Content Range	0–25% X 1%
Overall Dimensions	375 X 390 X 130 (L X W X D) mm
Net Weight	5.5 kg

2. Rapid Moisture Meter – 0–50%

Compliance Standards

IS:2720 (PART II), ASTM D4944, AASHTO T217, BS 812-109

Details

The unit consists of a pressure vessel with clamp for sealing cap, rubber sealing gasket, pressure gauge calibrated in percentage moisture content 0-50% X 1% on the wet weight basis, a counter poised balance for weighing sample, a scoop for measuring carbide reagent, a bottle of reagent, one cleaning brush and a set of 4 steel balls (3 OF 12.5MM AND 1 OF

Corporate Office:

A-1301, BVR Ek, Opp. Hotel Inder Residency,
Near Gujarat College, Ellisbridge,
Ahmedabad-380 006.
Phone : (079) 6621 1234
Website : www.eieinstruments.com

Calibration Laboratory:

Survey No. 1098, Nr. Mahakali Temple,
Opp. Primary School, Village Kubadthal,
Tal-Daskroi, Ahmedabad-382 433.
Phone: (079)-3520 8300
E-mail : info@eieinstruments.com

GST TIN No. : 24AABCE4018L1ZU
I.T. PAN No. : AABCE 4018L
IEC No. : 0813026598 Dt. 04/03/2014
CIN No. : U29199GJ2004PTC045078
TAN No. : AHME00572F



EIE INSTRUMENTS PVT. LTD.
(An ISO 9001:2015 Certified Company)

25MM DIA) for thorough mixing. Complete in highly polished wooden carrying case with handle.

Corporate Office:

A-1301, BVR Ek, Opp. Hotel Inder Residency,
Near Gujarat College, Ellisbridge,
Ahmedabad-380 006.
Phone : (079) 6621 1234
Website : www.eieinstruments.com

Calibration Laboratory:

Survey No. 1098, Nr. Mahakali Temple,
Opp. Primary School, Village Kubadthal,
Tal-Daskroi, Ahmedabad-382 433.
Phone: (079)-3520 8300
E-mail : info@eieinstruments.com

GST TIN No. : 24AABCE4018L1ZU
I.T. PAN No. : AABCE 4018L
IEC No. : 0813026598 Dt. 04/03/2014
CIN No. : U29199GJ2004PTC045078
TAN No. : AHME00572F