




EIE INSTRUMENTS PVT. LTD.
(An ISO 9001:2015 Certified Company)

VISUAL ASSESSMENT CABINET-VIEWING CABINET-BIS 13630 (PART XI), ISO 10545-7



Compliance Standards:

ISO 10545 (Part-7), BIS 13630 (Part-XI)

Details of the apparatus:

1. The apparatus consists of a viewing box equipped with fluorescent lighting
2. **Fluorescent light offers color temperature of 6000 K to 6500 K**
3. The lightings are placed vertically above the surface to be observed providing 300 lux illuminance.
4. **The dimensions of the box is 61 x 61 x 61 cm**
5. **The inner faces are painted a neutral grey.**
6. Light source is screened to avoid direct viewing.
7. Lux meter available at extra cost

Corporate Office:

A-1301, BVR Ek, Opp. Hotel Inder Residency,
Near Gujarat College, Ellisbridge,
Ahmedabad-380 006.
Phone : (079) 6621 1234
Website : www.eieinstruments.com

Calibration Laboratory:

Survey No. 1098, Nr. Mahakali Temple,
Opp. Primary School, Village Kubadthal,
Tal-Daskroi, Ahmedabad-382 433.
Phone: (079)-3520 8300
E-mail : info@eieinstruments.com

GST TIN No. : 24AABCE4018L1ZU
I.T. PAN No. : AABCE 4018L
IEC No. : 0813026598 Dt. 04/03/2014
CIN No. : U29199GJ2004PTC045078
TAN No. : AHME00572F