



EIE INSTRUMENTS PVT. LTD.
(An ISO 9001:2015 Certified Company)

V FUNNEL TEST APPARATUS



Compliance standard

IS 1199:PART 6/EN 12350-9: 2010

Application of V-Funnel test method for SCC :

The V- Funnel flow time is the period in which, a defined volume of SCC (self-consolidating concrete) needs to pass through a narrow opening and gives an indication of the filling ability of SCC, provided that blocking and or segregation do not take place. The flow time of the V-Funnel test is to some degree related to the plastic viscosity.

Construction Details :

The apparatus consists of a V-shaped funnel, which is supported on a Mild steel rigid frame. The frame and the funnel are spray painted or powder coated in attractive shades..

Essential Accessories at extra cost as per following :

1. Scoop
2. Ruler
3. Stopwatch
4. Trowel

Note : Also available in all S.S. or Acrylic construction.

Technical Specification

- Overall Dimensions : 570 X 570 X 1120(L X W X D) mm
- Weight (Approx) : 16 kg

Corporate Office:

A-1301, BVR Ek, Opp. Hotel Inder Residency,
Near Gujarat College, Ellisbridge,
Ahmedabad-380 006.
Phone : (079) 6621 1234
Website : www.eieinstruments.com

Calibration Laboratory:

Survey No. 1098, Nr. Mahakali Temple,
Opp. Primary School, Village Kubadthal,
Tal-Daskroi, Ahmedabad-382 433.
Phone: (079)-3520 8300
E-mail : info@eieinstruments.com

GST TIN No. : 24AABCE4018L1ZU
I.T. PAN No. : AABCE 4018L
IEC No. : 0813026598 Dt. 04/03/2014
CIN No. : U29199GJ2004PTC045078
TAN No. : AHME00572F