



EIE INSTRUMENTS PVT. LTD.
(An ISO 9001:2015 Certified Company)

VISCOMETER CLEANING AND DRYING APPARATUS- ASTM D445



Corporate Office:

A-1301, BVR Ek, Opp. Hotel Inder Residency,
Near Gujarat College, Ellisbridge,
Ahmedabad-380 006.
Phone : (079) 6621 1234
Website : www.eieinstruments.com

Calibration Laboratory:

Survey No. 1098, Nr. Mahakali Temple,
Opp. Primary School, Village Kubadthal,
Tal-Daskroi, Ahmedabad-382 433.
Phone: (079)-3520 8300
E-mail : info@eieinstruments.com

GST TIN No. : 24AABCE4018L1ZU
I.T. PAN No. : AABCE 4018L
IEC No. : 0813026598 Dt. 04/03/2014
CIN No. : U29199GJ2004PTC045078
TAN No. : AHME00572F



Compliance Standards:

ASTM D445

Salient Features:

- Cleans up to 6 tubes simultaneously
- Heated solvent vapor process
- 5 to 15 minute cleaning time
- Low solvent loss

Details on functioning of the equipment:

- The Viscometer Tube Cleaner is a fast and efficient instrument that will clean all current types of viscometer tube internally and externally.
- Up to six viscometer tubes can be suspended in the cleaning chamber.
- Solvent vapours, heated to 80°C, permeate the viscometer tubes, condensing on the internal and external surfaces, dissolving any residues from samples passed through the tube during a test.
- A water fed cooling coil at the top of the chamber causes the vapor to continually condense and return to the receiver, minimizing solvent loss.
- The use of a low temperature re-circulator (**Supplied at extra cost**) to chill the cooling water will improve the solvent recovery rate

Technical Specification

- Temperature Range : Ambient + 5 °C to 80 °C
- Temperature Accuracy : ± 0.5 °C
- Temperature Display : LED Digital display
- Temperature controller : Microprocessor based DIGITAL PID Controller
- Temperature Resolution : 0.1 °C
- RELEVANT STANDARD : ASTM D445
- Temperature sensor : PT-100 Sensor
- Heater Rating : 1500 watts
- Weight (Approx.) : 26.500 kg
- Power Supply : 230 Volts, 50 Hz, Single Phase, AC Supply

Corporate Office:

A-1301, BVR Ek, Opp. Hotel Inder Residency,
Near Gujarat College, Ellisbridge,
Ahmedabad-380 006.
Phone : (079) 6621 1234
Website : www.eieinstruments.com

Calibration Laboratory:

Survey No. 1098, Nr. Mahakali Temple,
Opp. Primary School, Village Kubadthal,
Tal-Daskroi, Ahmedabad-382 433.
Phone: (079)-3520 8300
E-mail : info@eieinstruments.com

GST TIN No. : 24AABCE4018L1ZU
I.T. PAN No. : AABCE 4018L
IEC No. : 0813026598 Dt. 04/03/2014
CIN No. : U29199GJ2004PTC045078
TAN No. : AHME00572F